


Verification of Conformity

Applicant : BES Group/Yantai United Ozonetec Corporation
Address : 4F, No. 1 Wuzhishan Road, ETDZ, Yantai Shandong Province, China
264006
Manufacturer : BES Group/Yantai United Ozonetec Corporation
Address : 4F, No. 1 Wuzhishan Road, ETDZ, Yantai Shandong Province, China
264006
Brand Name : Biosue
Product : Aqueous Ozone Dental Unit Waterline Cleaning System - DWCS160
Model No. : EOS7162-D
Rating : Battery 3.7Vdc
Charging by DC 5V from adapter
Test Report No. : DNT2509050151E9830-11927
Directive : Electromagnetic Compatibility Directive (2014/30/EU)
Test Standard(s) : EN IEC 55014-1:2021
EN IEC 55014-2:2021
EN IEC 61000-3-2:2019+A1:2021
EN 61000-3-3:2013+A1:2019+A2:2021

This Verification of Conformity is based on evaluation of a sample of the above mentioned product. Technical Report and documentation are at the applicant's disposal. This is to verify that the tested sample is in conformity with all revision of Electromagnetic Compatibility Directive, in its latest amended version, referred to as the EMC Directive. This document does not imply assessment of the product's production and does not permit the use of DNT mark. The holder of this document is authorized to use this document in connection with the EC declaration of conformity.





Manager

Date
2025/09/19